

ANSWER TO QUESTION - PART B #1

The standard uncertainty of our centigram balance is ± 0.01 gram. Therefore all mass readings should be to TWO decimal places.

The student is incorrect, because her reading (28.2 grams) only goes to the first decimal place.